

NABL ACCREDITED LABORATORY FOR CALIBRATION OF MEDICAL EQUIPMENT

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	VENTILATOR	TIDAL VOLUME, RESPIRATORY RATE, MEAN AIRWAY PRESSURE, POSITIVE END EXPIRATORY PRESSURE (PEEP), OXYGEN CONCENTRATION	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	VENTILATOR	TIDAL VOLUME, FIO ₂ , BREATH RATE, I:E RATIO, PEEP	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	VENTILATOR	VOLUME ACCURACY, RESPIRATION RATE, INSPIRATORY EXPIRATORY TIME, PRESSURE ACCURACY (PIP), PEEP, OXYGEN PERCENTAGE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

4	VENTILATOR	RESPIRATION RATE, VOLUME ACCURACY, PRESSURE ACCURACY OXYGEN PERCENTAGE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
5.	VENTILATOR	OXYGEN PERCENTAGE, VOLUME PRESSURE, PEEP, RESPIRATION RATE,	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080- 27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA
6.	VENTILATOR	FIO ₂ , PRESSURE (PEEP, PIP, MAP), RESPIRATION RATE I:E RATIO, VOLUME	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN
7.	VENTILATOR	INSPIRATORY EXPIRATORY TIME , PRESSURE (PEEP), OXYGEN PERCENTAGE RESPIRATORY RATE, PRESSURE ACCURACY, VOLUME ACCURACY	NORTHLAB INDIA PVT LTD	PLOT NO 1 SV NAGAR PERUMALPATTU, CHENNAI.	TAMIL NADU	8939923966	rajasekar@northlab.in	K RAJASEKARAN

8	VENTILATOR	VOLUME, PEEP, RESPIRATION RATE,	KIRLOSKAR TECHNOLOGIES (P) LTD. (QUALITY ASSESSMENT SERVICES)	B-58, 1ST FLOOR DEFENCE COLONY, BHISHAM PITAMAH MARG, NEW DELHI	DELHI	7018270274	gianender.verma@ kirloskarmedical.com	MR. GIANENDER SINGH VERMA
9	VENTILATOR	INSPIRATORY VOLUME, EXPIRATORY VOLUME, PEEP, OXYGEN CONCENTRATION, MINUTE VOLUME, ENCLOSURE LEAKAGE	TRIMED SOLUTIONS (INDIA) PRIVATE LIMITED	EXPRESS CENTRE,5TH FLOOR,PLOT NO.2, H.NO.6-3-1186,LANE NEXT TO ITC KAKATIYA HOTEL, BEGUMPET, HYDERABAD &NDASH; 500 016, HYDrabad	TELANGANA	8019779199	quality@trimedolutions.in	MR.BUNA ANIL

S. NO	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	PULSE OXIMETER	HEART RATE SpO2	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	PULSE OXIMETER	HEART RATE SpO2	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	PULSE OXIMETER	HEART RATE SpO2	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KARNATAKA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	PULSE OXIMETER	HEART RATE SpO2	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
5	PULSE OXIMETER	PULSE RATE SpO2	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN
6	PULSE OXIMETER	PULSE RATE SpO2	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080-27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA
7	Monitoring Unit	SpO2	QUALITY CERTIFICATION & LABORATORIES	PLOT NO-10 DSIDC SCHEME-3 1ST FLOOR, OKHLA INDL. AREA, PHASE-II, DELHI	DELHI	9818009007	v_devgun@hotmail.com	MR.VIVEK DEVGUN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	DISCHARGE EQUIPMENT/DEVICES	FLOW - OXYGEN CONCENTRATORS	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	MR.VENKATRAMAN
2	DISCHARGE EQUIPMENT/ DEVICES	OXYGEN CONCENTRATION	QUALITY CERTIFICATION & LABORATORIES	PLOT NO-10 DSIDC SCHEME-3 1ST FLOOR, OKHLA INDL. AREA, PHASE-II, DELHI	DELHI	9818009007	v_devgun@hotmail.com	MR.VIVEK DEVGUN
3	DISCHARGE EQUIPMENT/ DEVICES	VOLUME, PEEP, RESPIRATION RATE-OXYGEN CONCENTRATION MACHINE, MEDICAL OXYGEN FLOW METER WITH HUMIDIFIER, MEDICAL OXYGEN PRESSURE GAUGES AND TRANSMITTERS	KIRLOSKAR TECHNOLOGIES (P) LTD. (QUALITY ASSESSMENT SERVICES)	B-58, 1ST FLOOR DEFENCE COLONY, BHISHAM PITAMAH MARG, NEW DELHI	DELHI	7018270274	gianender.verma@kirloskarmedical.com	MR. GIANENDER SINGH VERMA

4.	DISCHARGE EQUIPMENT/ DEVICES	FLOW – FLOW METER	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080-27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA
5.	DISCHARGE EQUIPMENT/ DEVICES	Spo2 ACCURACY - PATIENT MONITOR	NORTHLAB INDIA PVT LTD	PLOT NO 1 SV NAGAR PERUMALPATTU, CHENNAI	TAMIL NADU	8939923966	rajasekar@northlab.in	MR.K RAJASEKARAN
6.	DISCHARGE EQUIPMENT/ DEVICES	PATIENT MONITOR - SATURATION PERCENTAGE OF OXYGEN(SPO2)	TRIMED SOLUTIONS (INDIA) PRIVATE LIMITED	EXPRESS CENTRE,5TH FLOOR,PLOT NO.2, H.NO.6-3-1186,LANE NEXT TO ITC KAKATIYA HOTEL, BEGUMPET, HYDERABAD &NDASH; 500 016, HYDrabad	TELANGANA	8019779199	quality@trimedolutions.in	MR.BUNA ANIL

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP, SPO ₂ , INVASIVE PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP LEAK TEST, NIBP DYNAMIC, SPO ₂	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	PATIENT MONITOR	ECG, NIBP, SPO ₂ , TEMPERATURE, RESPIRATION	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP LEAKAGE TEST, NIBP TEST, NIBP PRESSURE RELIEF, INVASIVE BLOOD PRESSURE, SPO ₂ , HEART RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

5.	PATIENT MONITOR	IABP TEMPERATURE RESPIRATION RATE HEART RATE NIBP DYNAMIC NIBP LEAK TEST	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA
6	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP DYNAMIC, IABP SPO ₂ NIBP LEAKAGE TEST, TEMPERATURE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI-	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
7.	PATIENT MONITOR	HEART RATE RESPIRATION RATE, NIBP DYNAMIC	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080- 27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	INFUSION PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	INFUSION PUMP	FLOW RATE ACCURACY, VOLUME ACCURACY, OCCLUSION DETECTION PRESSURE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI

6	INFUSION PUMP	FLOW, VOLUME, OCCLUSION	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080-27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA
7.	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION DETECTION PRESSURE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K

4	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	SYRINGE PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
6	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION	ROOTS METROLOGY & TESTING LABORATORY	PLOT NO. 156, BOMMASANDRA INDUSTRIAL AREA, BOMMASANDRA, JIGANI LINK ROAD, ANEKAL TALUK, BANGALORE	KARNATAKA	080-27832261, 07022968854	blrmetlab@rootsemail.com	Mr. J. SIVA
7.	SYRINGE PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, DISCHARGE TIME, SYNCHRONIZER OPERATION	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, ENERGY AFTER 10 CONSECUTIVE DISCHARGES, TIMELY CHARGING	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	DEFIBRILLATOR	ECG, ENERGY, SYNC TEST	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM,		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, OUTPUT ENERGY AT MAXIMUM SETTING FOR 10CHg CYCLE (AT BATTERY POWER), CHARGE TIME AFTER 10 DISCHARGE CYCLES, ENERGY AFTER 60 SEC OF FULL CHARGE AT MAX ENERGY, SYNCHRONIZER OPERATION, NIBP TEST, O ₂ SATURATION SPO ₂ (HEART RATE), PACER AMPLITUDE, PACER RATE ACCURACY	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

5	DEFIBRILLATOR	OUTPUT ENERGY, DISCHARGE TIME (POSITIVE / NEGATIVE PULSE WIDTH), HEART RATE, SYNCHRONI ZATION TEST- CHARGE TIME	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA
6	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, DISCHARGE TIME, SYNCHRONIZATION TEST, PACER OUTPUT PACER RATE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
7.	DEFIBRILLATOR	OUTPUT, ECG RATE, ENERGY, PACER OUTPUT PACER RATE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	SUCTION PUMP/ ASPIRATOR	VACUUM (PRESSURE)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	SUCTION PUMP	PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO- ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	SUCTION PUMP	VACUUM	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080- 23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
4	SUCTION PUMP	VACUUM(PRESSURE)	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BP APPARATUS	LEAK TEST NIBP (STATIC PRESSURE)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	BP APPARATUS	NIBP LEAK TEST	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	BP APPARATUS	PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	BP APPARATUS	SPHYGMOMANOMETER: LEAK TEST GAUGE ZERO PRESSURE ACCURACY DIGITAL BP APPARATUS: LEAK TEST PRESSURE ACCURACY HEART RATE ACCURACY	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAK A	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5.	BP APPARATUS	NIBP- STATIC PRESSURE NIBP LEAK TEST	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925 9841392570	pmathane@godrej.com, kp@godrej.com	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
6.	BP APPARATUS	PRESSURE (STATIC, DYNAMIC, RELIFE)	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	ECG MACHINE	HEART RATE AMPLITUDE	APEX QUALITY CERTIFICATION SERVICES PVT LTD	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	ECG MACHINE	HEART RATE AMPLITUDE	CYRIX HEALTH CARE PRIVATE LTD, CALIBRATION LABORATORY DIVISION	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	ECG MACHINE	HEART RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO- ORPORATION, TRIVANDRUM TALUK,		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	ECG MACHINE	HEART RATE, AMPLITUDE,	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	ECG MACHINE	HEART RATE, AMPLITUDE PATIENT LEAD LEAKAGE CURRENT, CHASIS LEAKAGE, GROUND WIRE RESISTANCE, INSULATION RESISTANCE, PATIENT LEAKAGE,	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF- 85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA
6	ECG MACHINE	HEART RATE, AMPLITUDE,	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
7.	ECG MACHINE	HEART RATE, OUTPUT	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR,CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	NEBULIZER	FLOW RATE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	NEBULIZER	FLOW RATE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	NEBULIZER	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	NEBULIZER	FLOW RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	NEBULIZER	FLOW RATE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341 , 0226796592 5, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	RADIANT WARMER AND INFANT INCUBATOR	TEMPERATURE OF WARMER, HUMIDITY IN WARMER, AIR FLOW, SOUND LEVEL	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	RADIANT WARMER	TEMPERATURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	RADIANT WARMER	TEMPERATURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
4	RADIANT WARMER , PATIENT WARMER	TEMPERATURE	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA
5.	RADIANT WARMER	TEMPERATURE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
6.	WARMER	TEMPERATURE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	AUTOCLAVE	TEMPERATURE PRESSURE GAUGE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080- 23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
2	AUTOCLAVE	TEMPERATURE PRESSURE	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA
3	AUTOCLAVE	TEMPERATURE PRESSURE GAUGE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
4	AUTOCLAVE	TEMPERATURE PRESSURE GAUGE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BOYLES APPARATUS	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
2	BOYLES APPARATUS	FLOW RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
3	BOYLES APPARATUS	FLOW, PRESSURE	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN

S. NO	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BLOOD GAS ANALYZER, SEMI AUTO ANALYSER, HEMATOLOGY ANALYZER, O.T. TABLE,	GROUND WIRE RESISTANCE, CHASIS LEAKAGE CURRENT, PATIENT LEAKAGE CURRENT, PATIENT LEAD LEAKAGE CURRENT, ISOLATION TEST (MAINS ON APPLIED PARTS) INSULATION TEST (OPTIONAL 500V)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	BLOOD GAS ANALYZER, SEMI AUTO ANALYSER, HEMATOLOGY ANALYZER, O.T. TABLE,	GROUND WIRE RESISTANCE, CHASIS LEAKAGE CURRENT, PATIENT LEAKAGE CURRENT, PATIENT LEAD LEAKAGE CURRENT, INSULATION TEST (OPTIONAL 500V)	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI-	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI
3	FLOW METER	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	FLOW METER	FLOW RATE	GODREJ & BOYCE MFG.CO.LTD	LAWKIM MOTORS GROUP, PLANT 18B, FIRST FLOOR, PIROJSHANAGAR, VIKHROLI, MUMBAI-	MAHARASHTRA	9967911341, 02267965925, 9841392570	9967911341, 02267965925, 9841392570	Mr.PRASHANT MATHANE Ms.K.PARAMESWARI

5	FLOW METER	FLOW	BIO VISION MEDICAL SYSTEMS	NO 49&50, SECOND FLOOR, SAMUVEL NAGAR, CHENNAI	TAMIL NADU	7338888671	venkat.biomed@gmail.com	Mr.VENKATRAMAN
	BiPAP	RESPIRATORY RATE, PRESSURE (IPAP, EPAP)						
6	BiPAP	IPAP, EPAP, RR, Ti	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080- 23362266	mail@tepl.info	MR. G.K. SURESH KUMAR